

## Certificate of Calibration

**Detail :** DRY BATH ID No. : DRY-LAB-01  
Manufacturer : BIOLOGIC Model : HQ210 Serial No. : 65150394007-1  
**Department :** โรงพยาบาลหนองหิน Section : LAB  
หนองหิน จังหวัดเลย  
**Room Ambient :** Temperature : 23 °C ( 15-25 °C) Humidity : 50 %RH ( 20-85 %RH )  
**Cal Date :** 2/10/2022 **Approved Date** 3/10/2022

**Measurement Method :** Follow the dry bath calibration method of W๑-CAL-TMA-๐๑. By using measuring instruments that have been calibrated by a standardized agency and calibrated under specified environmental conditions.

### Standard Used :

Equipment	Manufacture	Model	Serial No.	Certification No.	Cal Date
Temp - Humidity Meter	KIMO	HQ210	2P150601383	21H1601	7/16/2021

### Measurement Results :

UUC Setting (°C)	STD Reading (°C)	Error (°C)	Uncertainty (+/-°C)	Accept value (+/-°C)
55.600	53.500	2.100	2.317	2.00

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , Providing a level of confidence of approximately 95%

Cal by : Mr.Kitipong Saengkaew

Approved by :

*Pht. T.*  
( Mr.Phatchara Temrat )  
Biomedical Engineer

**Comment :** This report certifies the calibrate results as of the date, the location and the conditions of calibrate only.