



Certificate of Calibration

Detail : REFRIG.THERMOMETER ID No. : TMR-PHA-07
 Manufacturer : - Model : - Serial No. : -
Department : โรงพยาบาลหนองหิน Section : PHA
 หนองหิน จังหวัดเลย
Room Ambient : Temperature : 23 °C (15-25 °C) Humidity : 50 %RH (20-85 %RH)
Cal Date : 2/10/2022 **Approved Date** 3/10/2022

Measurement Method : Follow the thermometer calibration method of W๑-CAL-TM-๐๑. By using measuring instruments that have been calibrated by a standardized agency and calibrated under specified environmental conditions.

Standard Used :

Equipment	Manufacture	Model	Serial No.	Certification No.	Cal Date
Temp - Humidity Meter	KIMO	HQ210	2P150601383	21H1601	7/16/2021

Measurement Results :

POSITION	STD Setting (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)	Accept value (+/-°C)
	4.000	3.200	-0.800	0.118	1.00

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, Providing a level of confidence of approximately 95%

Cal by : Mr. Suthad Pornpipat

Approved by :

Phat. T.

(Mr.Phatchara Temrat)
Biomedical Engineer

Comment : This report certifies the calibrate results as of the date, the location and the conditions of calibrate only.