



## Certificate of Calibration

**Detail :** PATIENT THERMOMETER ID No. : TMP-CK-01  
 Manufacturer : - Model : - Serial No. : -  
**Department :** โรงพยาบาลหนองหิน Section : รพ.สต.เฉลิมพระเกียรติ  
 หนองหิน จังหวัดเลย  
**Room Ambient :** Temperature : 23 °C ( 15-25 °C ) Humidity : 50 %RH ( 20-85 %RH )  
**Cal Date :** 2/10/2022 **Approved Date** 3/10/2022

**Measurement Method :** Follow the thermometer calibration method of W๑-CAL-TM-๐๑. By using measuring instruments that have been calibrated by a standardized agency and calibrated under specified environmental conditions.

**Standard Used :**

Equipment	Manufacture	Model	Serial No.	Certification No.	Cal Date
Temp - Humidity Meter	KIMO	HQ210	2P150601383	21H1601	7/16/2021

**Measurement Results :**

POSITION	STD Setting (°C)	UUC Reading (°C)	Error (°C)	Uncertainty (°C)	Accept value (+/-°C)
	37.000	37.000	0.000	0.118	0.50

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , Providing a level of confidence of approximately 95%

Cal by : Mr. Suthad Pornpipat

Approved by :

*Pht. T.*

( Mr.Phatchara Temrat )

Biomedical Engineer

**Comment :** This report certifies the calibrate results as of the date, the location and the conditions of calibrate only.