

## Certificate of Calibration

**Detail :** FETAL MONITOR ID No. : FTM-LR-01  
 Manufacturer : FAZZINI Model : FM5000D Serial No. : 54140402032

**Department :** โรงพยาบาลหนองหิน Section : LR  
 หนองหิน จังหวัดเลย

**Room Ambient :** Temperature : 23 °C ( 15-25 °C ) Humidity : 50 %RH ( 20-85 %RH )

**Cal Date :** 2/10/2022 **Approved Date** 3/10/2022

**Measurement Method :** Follow the Fatal monitor calibration method of W<sub>6</sub>-CAL-FMT-๐๑. By using measuring instruments that have been calibrated by a standardized agency and calibrated under specified environmental conditions.

### Standard Used :

Equipment	Manufacture	Model	Serial No.	Certification No.	Cal Date
Fetal Monitor Analyzer	Fluke	PS320	2093004	21MD882	7/20/2021

### Measurement Results :

Fetal Heart Rate	STD Setting (BPM)	UUC Reading (BPM)	Error (BPM)	Uncertainty (+/-BPM)	Accept value (+/-BPM)
	60.000	60.000	0.000	0.089	1.00
	90.000	90.000	0.000	0.089	1.00
	120.000	120.000	0.000	0.089	1.00
	150.000	150.000	0.000	0.089	1.00
	180.000	180.000	0.000	0.089	1.00
	210.000	210.000	0.000	0.089	1.00

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor  $k = 2$ , Providing a level of confidence of approximately 95%

Cal by : Mr. Suthad Pornpipat

Approved by :

P.W.T.

( Mr.Phatchara Temrat )

Biomedical Engineer

**Comment :** This report certifies the calibrate results as of the date, the location and the conditions of calibrate only.