

Certificate of Calibration

Detail : HEMATOCRIT ID No. : HEM-IPD-01
Manufacturer : BOECO Model : H-240 Serial No. : 0002256-02-00

Department : โรงพยาบาลหนองหิน Section : IPD
หนองหิน จังหวัดเลย

Room Ambient : Temperature : 23°C (15-25°C) Humidity : 50 %RH (20-85 %RH)

Cal Date : 2/10/2022 **Approved Date** 3/10/2022

Measurement Method : Follow the centrifuge calibration method of W๑-CAL-HEM,CEN,SER.RTT-๐๑. By using measuring instruments that have been calibrated by a standardized agency and calibrated under specified environmental conditions.

Standard Used :

Equipment	Manufacture	Model	Serial No.	Certification No.	Cal Date
Techo Meter	Testo	470	1304078	21-E2187	7/13/2021

Measurement Results :

Speed	UUC Setting (RPM)	STD Reading (RPM)	Error (RPM)	Uncertainty (+/-RPM)	Accept value (+/-RPM)
	12000.00	11985.00	15.00	0.58	1200.00

The reported uncertainty is based on a standard uncertainty multiplied by a coverage factor $k = 2$, Providing a level of confidence of approximately 95%

Cal by : Mr. Suthad Pornpipat

Approved by :

Ph. T.

(Mr.Phatchara Temrat)
Biomedical Engineer

Comment : This report certifies the Calibrate results as of the date, the location and the conditions of calibrate only.